



Your PDF Guides

You can read the recommendations in the user guide, the technical guide or the installation guide for HANNA INSTRUMENTS HI 98312. You'll find the answers to all your questions on the HANNA INSTRUMENTS HI 98312 in the user manual (information, specifications, safety advice, size, accessories, etc.). Detailed instructions for use are in the User's Guide.

User manual HANNA INSTRUMENTS HI 98312 User guide HANNA INSTRUMENTS HI 98312 Operating instructions HANNA INSTRUMENTS HI 98312 Instructions for use HANNA INSTRUMENTS HI 98312 Instruction manual HANNA INSTRUMENTS HI 98312

Instruction Manual HI 98311 • HI 98312 Waterproof EC/TDS/Temperature Testers



WARRANTY

HI 98311 and HI 98312 are warranted for one year against defects in workmanship and materials when used for their intended purpose and maintained according to instructions. The probe is warranted for a period of six months. This warranty is limited to repair or replacement free of charge.

Damages due to accident, misuse, tampering or lack of prescribed maintenance are not covered. If service is required, contact the dealer from whom you purchased the instrument. If under warranty, report the model number, date of purchase, serial number and the nature of the failure. If the repair is not covered by the warranty, you will be notified of the charges incurred. If the instrument is to be returned to Hanna Instruments, first obtain a Returned Goods Authorization Number from the Customer Service department and then send it with shipment costs prepaid. When shipping any instrument, make sure it is properly packaged for complete protection.

All rights are reserved. Reproduction in whole or in part is prohibited without the written consent of the copyright owner, Hanna Instruments.

Hanna Instruments reserves the right to modify the design, construction and appearance of a product without advance notice.

Dear Customer,
Thank you for choosing a Hanna product. This manual will provide you with the necessary information for a correct operation. Please read it carefully before using the meter.
If you need additional technical information, do not hesitate to e-mail us at tech@hannainst.com.

These instruments are in compliance with the CE directives.

PRELIMINARY EXAMINATION

Remove the instrument from the packing material and examine it carefully. If any damage has occurred during shipment, immediately notify your Dealer or the nearest Hanna Customer Service Center.

Each meter is supplied with:
• HI 73311 EC/TDS probe
• HI 73128 probe removal tool
• Batteries (4 x 1.5V) and instructions

Note: Conserve all packing material until the instrument has been observed to function correctly. Any defective item must be returned in its original packaging.

US DESIGN PATENT
D462,024

GENERAL DESCRIPTION

HI 98311 and HI 98312 are waterproof EC/TDS and temperature meters. The housing has been completely sealed against humidity and designed to float. All EC/TDS readings are automatically temperature compensated (ATC), and temperature values can be displayed in °C or °F units.


The EC/TDS conversion factor (CONV) is selectable by the user, as well as the temperature compensation coefficient (BETA).

The meters can be calibrated at one point. Measurements are highly accurate with a unique stability indicator right on the LCD.


A low battery symbol warns the user when the batteries are to be replaced. In addition the Battery Error Prevention System (BEP5) avoids erroneous readings caused by low voltage level by turning the meter off. The HI 73311 EC/TDS probe, supplied with the meter, is interchangeable and can be easily replaced by the user.

The stainless steel encapsulated temperature sensor is calibrated faster and more accurate temperature measurement and compensation.

FUNCTIONAL DESCRIPTION



1. Battery compartment
2. Liquid Crystal Display (LCD)
3. ON/OFF/MODE button
4. HI 73311 EC/TDS probe
5. Temperature sensor
6. SET/HOLD button



1. Automatic temperature compensation indicator
2. Stability indicator
3. Battery life percentage indicator
4. Low battery indicator
5. Secondary display
6. Primary display
7. Measuring units for primary display

SPECIFICATIONS

| | |
|--------------------------|---|
| Range | 0.0 to 60.0°C / 32.0 to 140.0°F 0 to 3999 µS/cm (HI 98311) 0.00 to 20.00 mS/cm (HI 98312) 0 to 2000 ppm (HI 98311) 0.00 to 10.00 ppt (HI 98312) |
| Resolution | 0.1°C / 0.1°F 1 µS/cm, 1 ppm (HI 98311) 0.01 mS/cm, 0.01 ppt (HI 98312) |
| Accuracy | ±0.5°C / ±1°F (@20°C/68°F) ±2%±1 (EC/TDS) |
| Typical EMC | ±0.5°C / ±1°F |
| Deviation | ±2%±1 (EC/TDS) |
| Temperature Compensation | Automatic with BETA: 0.0 to 2.4%/°C |
| Environment | 0 to 50°C (32 to 122°F), RH 100% |
| TDS Factor | 0.45 to 1.00 (CONV) |
| Calibration | Automatic, 1 point |
| Cal. Solutions | |
| HI 98311 | HI7031 (1413 µS/cm) HI7032 (1382 ppm, CONV=0.5) HI7042 (1500 ppm, CONV=0.7) HI7030 (12.88 mS/cm) |
| HI 98312 | HI7008 (6.44 ppt, CONV=0.5) HI7009 (9.02 ppt, CONV=0.7) |
| Probe (included) | HI 73311 EC/TDS probe |
| Battery Type/Use | 4 x 1.5V with BEP5 approx. 100 hours |
| Auto-off | After 8 minutes of non-use |
| Dimensions | 163 x 40 x 28 mm (6.4 x 1.6 x 1.0") |
| Weight | 100 g (3.5 oz) |

Recommendations for Users
Before using this product, make sure that it is entirely suitable for the environment in which it is used. Operation of this instrument in residential areas could cause unacceptable interference to radio and TV equipment.
Avoid touching the probe at all times.
Any variation introduced by the user to the supplied equipment may degrade the instrument's EMC performance. To avoid electrical shocks, do not use this instrument when voltages of the measurement surface exceed 24 Vac or 60 Vdc. To avoid damage or burns, do not perform any measurement in microwave ovens.

©19831183 07/05



[You're reading an excerpt. Click here to read official HANNA INSTRUMENTS HI 98312 user guide](http://yourpdfguides.com/dref/2866009)
<http://yourpdfguides.com/dref/2866009>

Manual abstract:

This manual will provide you with the necessary information for a correct operation. Please read it carefully before using the meter. If you need additional technical information, do not hesitate to e-mail us at tech@hannainst.com. @@@@ (EC/TDS) $\pm 0.5^{\circ}\text{C}$ / $\pm 1^{\circ}\text{F}$ $\pm 2\%$ f.s. (EC/TDS) Automatic, with =0.0 to 2.4%/°C 0 to 50°C (32 to 122°F); RH 100% 0.

45 to 1.00 (CONV) Automatic, 1 point HI7031 (1413 $\mu\text{S}/\text{cm}$) HI7032 (1382 ppm; CONV=0.5) HI70442 (1500 ppm; CONV=0.7) Typical EMC Deviation Temperature Compensation Environment TDS Factor Calibration Cal.Solutions HI 98311; HI 98312: HI7030 (12.88 mS/cm) HI70038 (6.44 ppt; CONV=0.5 or 9.02 ppt; CONV=0.7) www.

hannainst.com US DESIGN PATENT D462,024 1. 2. 3. 4. 5. 6. Battery compartment Liquid Crystal Display (LCD) ON/OFF/MODE button HI 73311 EC/TDS probe Temperature sensor Probe (included) Battery Auto-off Type/Life HI 73311 EC/TDS probe 4 x 1.5V with BEPS/approx. 100 hours After 8 minutes of non-use 163 x 40 x 26 mm (6.

4 x 1.6 x 1.0") 100 g (3.5 oz.) WARRANTY HI 98311 and HI 98312 are warranted for one year against defects in workmanship and materials when used for their intended purpose and maintained according to instructions. The probe is warranted for a period of six months. This warranty is limited to repair or replacement free of charge. Damages due to accident, misuse, tampering or lack of prescribed maintenance are not covered. If service is required, contact the dealer from whom you purchased the instrument. If under warranty, report the model number, date of purchase, serial number and the nature of the failure. If the repair is not covered by the warranty, you will be notified of the charges incurred. If the instrument is to be returned to Hanna Instruments, first obtain a Returned Goods Authorization Number from the Customer Service department and then send it with shipment costs prepaid. When shipping any instrument, make sure it is properly packaged for complete protection. All rights are reserved. Reproduction in whole or in part is prohibited without the written consent of the copyright owner, Hanna Instruments.

GENERAL DESCRIPTION HI 98311 and HI 98312 are waterproof EC/TDS and temperature meters. The housing has been completely sealed against humidity and designed to float. All EC/TDS readings are automatically temperature compensated (ATC), and temperature values can be displayed in °C or °F units. The EC/TDS conversion factor (CONV) is selectable by the user, as well as the temperature compensation coefficient (BETA). The meters can be calibrated at one point.

@@@@@ Automatic temperature compensation indicator 2. Stability indicator 3. Battery life percentage indicator 4. Low battery indicator 5. Secondary display 6. Primary display 7. @@@@ Avoid touching the probes at all times. @@@@ @@@% 100 BATT). Taking measurements Submerge the probe in the solution to be tested. Use plastic beakers to minimize any electromagnetic interferences.

Select either EC or TDS mode with the SET/HOLD button. @@@@ TEMP °C). @@@@ Press either button to return to the normal measuring mode.

@@ OFF will appear on the lower part of the display. Release the button. @@@@ @) After testing aggressive chemicals. c) Where high accuracy is required. d) At least once a month. To change the EC/TDS conversion factor (CONV) and the temperature compensation coefficient (BETA) - From measurement mode, press and hold the /MODE button until TEMP and the current temperature unit are displayed on the lower LCD (E.g. TEMP °C). - Press the /MODE button again to show the current conversion factor (E.g. 0.50 CONV).

· Press the SET/HOLD button to change the conversion factor. /MODE button to show the current · Press the temperature compensation coefficient (E.g. 2.1 BETA).

· Press the SET/HOLD button to change the the temperature compensation coefficient . · Press the /MODE button to return to the normal measuring mode.

Calibration procedure PROBE MAINTENANCE The EC/TDS probe can be easily replaced by using the supplied tool (HI 73128). Insert the tool into the probe cavity as shown below. BATTERY REPLACEMENT The meter displays the remaining battery percentage every time it is switched on. When the battery level is below 5%, the symbol on the bottom left of the LCD lights up to indicate a low battery condition. The batteries should be replaced soon. If the battery level is low enough to cause erroneous readings, the meter shows "0%" and the Battery Error Prevention System (BEPS) will automatically turn the meter off. To change the batteries, remove the 4 screws located on the top of the meter. Rotate the probe counterclockwise.

Pull the probe out by using the other side of the tool. Insert a new probe following the above instructions in reverse order. Once the top has been removed, carefully replace the 4 batteries located in the compartment while paying attention to their polarity. · From measurement mode, press and hold the /MODE button until CAL is displayed on the lower LCD. · Release the button and immerse the probe in the proper calibration solution: HI 7031 (1413 $\mu\text{S}/\text{cm}$) for HI 98311 and HI 7030 (12.88 mS/cm) for HI 98312. @@@@ @@@ The CAL symbol on the LCD means that the meter is calibrated. @@ The lower LCD will display ESC for 1 second and the meter will return to normal measurement mode. The CAL symbol on the LCD will disappear. The meter will be reset to the default calibration.

Replace the top, making sure that the gasket is properly seated in place, and tighten the screws to ensure a watertight seal. ACCESSORIES HI 73311 HI 73128 HI 70030P HI 70031P HI 70032P HI 70038P HI 70442P Replaceable EC/TDS probe Probe removal tool 12.88 mS/cm solution, 20 mL sachet (25 pcs) 1413 $\mu\text{S}/\text{cm}$ solution, 20 mL sachet (25 pcs) 1382 ppm solution, 20 mL sachet (25 pcs) 6.44 ppt solution, 20 mL sachet (25 pcs) 1500 ppm solution, 20 mL sachet (25 pcs).



[You're reading an excerpt. Click here to read official HANNA INSTRUMENTS HI 98312 user guide](http://yourpdfguides.com/dref/2866009)
<http://yourpdfguides.com/dref/2866009>